## **Automatic Cell Counter**

Suitable for counting mammalian and other cells

Specifications:

•

- Counting time : ≤ 30 Secs
- Cell Concentration Range: Sample chamber volume: 5 X 10<sup>4</sup> to 1 X 10<sup>7</sup> •
- Sample chamber volume: 10 ul
  - Cell diameter range: 5 to 60 microns
- Should be capable of storing data of 100 or more specimens
- Dilution calculation •
- Viability calculation
- Display of Cells image. •
- Printer port
- USB connectivity
- Trypan Blue compatibility